



BIOPHARM

Quality Laboratory Solutions and Chemicals

Phone: 870-389-6114
Phone: 800-443-8465
Fax: 870-389-6604
E-mail: support@bphchem.com

187 S. Tilley Rd. :: Hatfield, AR 71945 :: www.bphchem.com

CERTIFICATE OF CONFORMANCE/ CERTIFICATE OF ANALYSIS

This Biopharm Product has been manufactured according to a written procedure which provides complete traceability to the lot(s) of raw material(s) used in its preparation. In addition, our records document the methods used in the formulation and analysis of this product. These records are kept for a period of not less than one (1) year to furnish additional data should any questions arise at some future date.

PRODUCT NUMBER: BA8101

PRODUCT DESCRIPTION: Alkalinity Standard 1000 ppm

LOT NUMBER: J1025B

EXPIRE: 9/10/2027

REQUIREMENTS: 1000 +/- 3 ppm

ANALYSIS RESULTS: 1000 ppm

MANUFACTURE DATE: 9/10/2025

CHEMIST: M. P. S. J.

DATE: 9-17-2025

***THANK YOU FOR CHOOSING QUALITY PRODUCTS FROM
BIOPHARM***