



BIOPHARM

Quality Laboratory Solutions and Chemicals

Phone: 870-389-6114
Phone: 800-443-8465
Fax: 870-389-6604
E-mail: support@bphchem.com

187 S. Tilley Rd. • Hatfield, AR 71945 • www.bphchem.com

CERTIFICATE OF CONFORMANCE

This Biopharm Product has been manufactured according to a written procedure which provides complete traceability to the lot(s) of raw material(s) used in its preparation. In addition, our records document the methods used in the formulation and analysis of this product. These records are kept for a period of not less than one (1) year to furnish additional data should any question arise at some future date.

PRODUCT NUMBER: BC8158

PRODUCT DESCRIPTION: Conductivity Standard 750 ppm @ 25°

Manufactured with Sodium Chloride

LOT NUMBER: J1825A

EXPIRATION: 9/18/2027

ACCURACY: Manufactured to $\pm 0.25\%$

Accuracy at expiration $\pm 1\%$

ASSAY: 750 ppm NaCl @ 25°C

CHEMIST: *M. Tilly*

DATE: *9-18-2025*

CUSTOMER NAME: Kimball Electronics Group

CUSTOMER PRODUCT NUMBER: 4503722952

QUANTITY: 3 Pints

PURCHASE ORDER NUMBER: 4502459017

WORK ORDER NUMBER: 25-1713

NIST TRACABILITY: SRM 2201

***THANK YOU FOR CHOOSING QUALITY PRODUCTS FROM
BIOPHARM***