



BIOPHARM INC.

Quality Laboratory Solutions and Chemicals

Phone: 870-389-6114

Phone: 800-443-8465

E-mail: support@bphchem.com

Website: www.bphchem.com

187 S. Tilley Rd. ** Hatfield, AR 71945

CERTIFICATE OF ANALYSIS AND CALIBRATION

Certificate No.: CM0425A

Product No.: BB8034

Analysis/Lot No.: M0425A

Product Description: Buffer Reference Solution pH 9.18 $\pm$ 0.02 @ 25°C

Date of Manufacture: 12/04/2025

Expiration Date: 6/4/2027

Traceability: The calibration of this product is traceable to NIST or International Standards through the following standards and equipment used in its manufacture.

Instrument	ID Number	Calibration Expiration	Traceable Reference
pH Electrode	600308039042	N/A*	SM0425B
Thermometer	A9B0909	5/17/2027	C4515014

* Calibration performed at time of product manufacture

This Biopharm product has been manufactured according to a written batch record which provides complete traceability to the lot(s) of raw material(s) used in its preparation. In addition, our records document the methods used in the formulation and analysis of this product. These records are kept for a period of not less than one (1) year to furnish additional data should any question arise at some future date.

Accuracy: This product was manufactured using standards with an accuracy of $\pm$ 0.005 unit from 0 to 50°C. The accuracy at the expiration date will be within the specified uncertainty of the original value under normal conditions of storage and handling.

Analyzed by: Michael Tilley (Chemist)

Analysis Procedure: BSP012

Analysis Date: 12/04/2025

Procedure Deviations: None

Analysis Results: 9.18 $\pm$ 0.02 @ 25°C

Certifying Chemist: Michael Tilley (Chemist)

Calibration Status: Pass

Signature: 

Date: 12/4/2025