



# BIOPHARM

Quality Laboratory Solutions and Chemicals

Phone: 870-389-6114  
Phone: 800-443-8465  
Fax: 870-389-6604  
E-mail: support@bphchem.com

187 S. Tilley Rd. .. Hatfield, AR 71945 .. www.bphchem.com

## CERTIFICATE OF CONFORMANCE/ CERTIFICATE OF ANALYSIS

This Biopharm Product has been manufactured according to a written procedure which provides complete traceability to the lot(s) of raw material(s) used in its preparation. In addition, our records document the methods used in the formulation and analysis of this product. These records are kept for a period of not less than one (1) year to furnish additional data should any questions arise at some future date.

PRODUCT NUMBER: BE8867

PRODUCT DESCRIPTION: Electrode Storage Solution

LOT NUMBER: B0124C

REQUIREMENTS: pH 3.5 - pH 4.0

ANALYSIS RESULTS: pH 3.98

MANUFACTURE DATE: 2/1/2024

CHEMIST: Michael Kelly

DATE: 4-10-2024

**THANK YOU FOR CHOOSING QUALITY PRODUCTS FROM  
BIOPHARM**