Phone: 870-389-6114 Phone: 800-443-8465 Fax: 870-389-6604

E-mail: support@bphchem.com

187 S. Tilley Rd. - Hatfield, AR 71945 - www.bphchem.com

CERTIFICATE OF ANALYSIS AND CALIBRATION

Certificate No.:	CD1023A			
Product No:	BC4094-50,000	Analysis/Lot No.:	D1023A	
Product Description:	Conductivity Standard 50,000 μMho/cm @ 25°C			
Date of Manufacture:	4/10/2023	Expiration Date: 4/1/2025		
Traceability: The calibration of this product is traceable to NIST or International Standards through the following standards and equipment used in its manufacture.				
	Instrument	ID Number Calibi	ation Expiration	Traceable Reference
	Conductivity Meter & Cell Thermometer	05D-97C-K10 A9B0909	6/21/2023 5/7/2024	SF2122C C1427018
This Biopharm Product has been manufactured according to a written batch record which provides complete traceability to the lot(s) of raw material(s) used in its preparation. In addition, our records document the methods used in the formulation and analysis of this product. These records are kept for a period of not less than one (1) year to furnish additional data should any question arise at some future date. Accuracy: This product was manufactured using standards with an accuracy of 0.25%. The accuracy at the expiration date will be within 1% of the original value under normal conditions of storage and handling.				
	The reported uncertainty (U) is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" and corresponds to a coverage probability of approximately 95% ($k=2$).			
Analyzed by:	Michael Tilley (Chemist)	Analysis Procedur	·e:	BSP0020
Analysis Date:	4/10/2023	Procedure Deviati	ons:	None
Analysis Results:	50,000 μMho/cm ; U = (+/-) 220 uMho/cm @ 25.0°C			
Certifying Chemist:	Michael Tilley (Chemist)	Calibration Status	s:	Pass
Signature:	Michael Till,	Date: Ø4	-17-2023	