CF2222A

Certificate No.:

Phone: 870-389-6114 Phone: 800-443-8465 Fax: 870-389-6604

E-mail: support@bphchem.com

187 S. Tilley Rd. - Hatfield, AR 71945 - www.bphchem.com

CERTIFICATE OF ANALYSIS AND CALIBRATION

Product No:	BC4094 – 100,000	Analysis/Lot No.:	F2222A	
Product Description:	Conductivity Standard 10	•		
Date of Manufacture:	6/22/2022	Expiration Date:	6/1/2024	
Traceability: The calibration of this product is traceable to NIST SRM 999 and International Standards through the following standards and equipment used in its manufacture.				
		Calibr	ation	Traceable
	Instrument	ID Number Expir		Reference
	Conductivity Meter & Cell		/2023	SF2122A
	Thermometer	A9B0909 5/3/3	2024	C1427018
formulation and analysis of	erial(s) used in its preparation. In addition, our records document the methods used in the of this product. These records are kept for a period of not less than one (1) year to furnish question arise at some future date. This product was manufactured using standards with an accuracy of 0.25%. The accuracy at the expiration date will be within 1% of the original value under normal conditions of storage and handling. The reported uncertainty (U) is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" and corresponds to a coverage probability of approximately 95% ($k=2$).			
Analyzed by:	Michael Tilley (Chemist)	Analysis Procedur	e: BS	SP0020
Analysis Date:	6/22/2022	Procedure Deviation	ons: No	one
Analysis Results:	100,000 μMho/cm; U = (+/-) 360 uMho/cm @ 25.0°C			
Certifying Chemist:	Michael Tilley (Chemist)	Calibration Status	: <u>Pa</u>	ass
Signature:	M. Lad Selle	Date: 6-29	7-22	