BIOPHARM INC.
Quality Laboratory Solutions and Chemicals
Phone: 870-389-6114
Phone: 800-443-8465
E-mail: support@bphchem.com
Website: www.bphchem.com
187 S. Tilley Rd. ** Hatfield, AR 71945

CERTIFICATE OF ANALYSIS AND CALIBRATION

Certificate No.: CD0325A

Product No: BB8870 **Analysis/Lot No.:** D0325A

Product Description: Buffer Reference Solution pH 10.00 ±0.02 @ 25°C

Date of Manufacture: 4/03/2025 **Expiration Date:** 10/3/2026

Traceability: The calibration of this product is traceable to NIST or International Standards through the following standards and equipment used in its manufacture.

Instrument	ID Number	Calibration Expiration	Traceable Reference
pH Electrode	600308039042	N/A*	SD0325A
Thermometer	A9B0909	5/17/2027	C4515014

* Calibration performed at time of product manufacture

This Biopharm product has been manufactured according to a written batch record which provides complete traceability to the lot(s) of raw material(s) used in its preparation. In addition, our records document the methods used in the formulation and analysis of this product. These records are kept for a period of not less than one (1) year to furnish additional data should any question arise at some future date.

Accuracy: This product was manufactured using standards with an accuracy of ±0.005 unit from 0 to 50°C. The accuracy at the expiration date will be within the specified uncertainty of the original value under normal conditions of storage and handling.

Analyzed by: Michael Tilley (Chemist) **Analysis Procedure:** BSP012

Analysis Date: 4/03/2025 **Procedure Deviations:** None

Analysis Results: 10.00 ±0.02 @ 25°C

Certifying Chemist: Michael Tilley (Chemist) **Calibration Status:** Pass

Signature: M. Tilley **Date:** 4/22/2025